



Calibration Certificate

Model: Focus S 70Serial Number: LLS081915769Certificate Number: LLS15769-20240730-EUCertification Date: 30-Jul-2024

Measurement Items Used, Traceable to National Standards

Laser Tracker

Model: Vantage ESerial No.: V20003706753Cert. No.: V06753-31102023Cert. Date: 31-Oct-2023

Reflectance Targets

Model: CalibLine Target90Serial No.: CTE01351Cert. No.: 23122104Cert. Date: 21-Dec-2023Model: CalibLine Target10Serial No.: CTE01346Cert. No.: 23122105Cert. Date: 21-Dec-2023Model: 0025-Target2Serial No.: CTE00711Cert. No.: 23122106Cert. Date: 21-Dec-2023

Calibration Results

3D Position Accuracy

Target 1	Target 2	Zone	Distance [m]	Uncertainty, k=2 [mm]	Scanner [m]	Deviation [mm]	Spec., k=1 [mm]	Spec., k=2 [mm]	Result
ZK6	ZK11	10 m	2,18214	0,20	2,18215	0,00	1,5	2,0	pass
ZK6	ZK8	10 m	2,59064	0,20	2,59003	-0,61	1,5	2,0	pass
ZK9	ZK10	25 m	5,15775	0,20	5,15752	-0,24	1,5	2,0	pass

Ranging Noise

Reflectance	Distance	Uncertainty, k=1 [mm]	Scanner	Specifications	Result
90%	10 m	0,067	0,08	0,30	pass
	25 m	0,067	0,11	0,30	pass
10%	10 m	0,067	0,18	0,40	pass
	25 m	0,067	0,33	0,50	pass
2%	10 m	0,067	0,48	1,30	pass
	25 m	0,067	1,00	2,00	pass

This certificate shall not be reproduced, except in full, without permission of FARO Technologies, Inc. It invalidates all other certificates generated before the certification date.
The results of this certificate relate only to the items calibrated or tested.
The calibration is done at FARO or FARO Scanner Production operations sites according to FARO test protocols integrating guidelines defined in the Joint Committee for Guides in Metrology guidance document JCGM 100:2008 - Evaluation of measurement data - Guide to the expression of uncertainty in measurement, and the requirements for traceability according to the International Vocabulary of Basic and General Terms in Metrology (VIM) and the National Institute of Standards and Technology (NIST).

Note: Data in this section is required only if amendments are made to the calibration certificate
Amendment to report: NO
Previous Certificate #: _____
Reason for Amendment: _____

Approved by: _____

Date: _____

FARO®

Revised: 12th July 2021 | © 2021 FARO | D3142

Authorization: Myrjam Kepura
Calibration TechnicianDate: 30-Jul-2024

FARO Europe GmbH - Lingwiesenstr. 11/2 - D 70825 Kornthal-Münchingen - Germany
Tel: +49 7150 9797-400 - Fax: +49 7150 9797-9400 - support.emea@faro.com - www.FARO.com